

DECLARATION OF CONFORMITY

MANUFACTURER: ANHUI DEEPBLUE MEDICAL TECHNOLOGY CO.,LTD.
4th Floor,D-1# Zone, Pearl Industrial Park, 106 Innovation Avenue,
High-Tech Development Zone , 230088 Hefei, Anhui People's
Republic of China

EUROPEAN REPRESENTATIVE: Luxus Lebenswelt GmbH
Kochstr. 1, 47877, Willich, Germany

PRODUCT: COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold)

Models: Swab

CLASSIFICATION: **OTHER**

EDMA CODE: 15 70 90 90 00

CONFORMITY ASSESSMENT ROUTE: Following the procedure relating to the EC Declaration of Conformity set out in Annex III of Directive 98/79/EC.

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCTS MEET THE PROVISIONS OF THE COUNCIL DIRECTIVE 98/79/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

THE MANUFACTURER IS EXCLUSIVELY RESPONSIBLE FOR THE DECLARATION OF CONFORMITY.

STANDARDS APPLIED: EN ISO 13485:2016
EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN 13612:
2002/AC:2002, EN ISO 23640:2015, EN 13641: 2002, EN ISO
15223-1: 2016, EN 13975:2003, EN ISO 14971:2012.

START OF CE-MARKING: 2020-07-31

PLACE, DATE OF ISSUE: HEFEI, 2021-03-29

SIGNATURE: **CHEN FENGLING**
GENERAL MANAGER

Chen Fengling



EC Declaration of Conformity
DOC-COVID-19 Ag- (F/0)